
Midmark M11 - v1.0.0
Total Cycles: 2131

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2022 02:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.5 F 0.0
2:00 179.1 F 3.7
4:00 197.7 F 8.4
6:00 225.5 F 8.9
8:00 228.8 F 8.6
10:00 240.0 F 12.1
12:00 239.8 F 10.9
14:00 237.1 F 9.0
16:00 251.2 F 15.8
18:00 262.9 F 22.5
20:00 272.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.3 F 28.6
2:00 272.2 F 28.8
3:00 271.8 F 28.4
4:00 272.4 F 29.0

Min 271.3 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 237.1 F 0.0
10:00 223.7 F 0.0
15:00 221.2 F 0.0
20:00 208.8 F 0.0
25:00 218.1 F 0.0
30:00 153.1 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 20:13
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:45
CYCLE COMPLETE

07/14/2022 03:40 PM
ACCEPTED: X
REJECTED:
Operator: U02