
Midmark M11 - v1.0.0
Total Cycles: 923

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2021 08:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.3 F 0.0
2:00 152.9 F 1.0
4:00 167.4 F 3.2
6:00 186.2 F 6.3
8:00 212.4 F 10.7
10:00 218.9 F 5.0
12:00 231.0 F 8.7
14:00 240.0 F 13.1
16:00 237.7 F 10.0
18:00 237.5 F 9.0
20:00 247.4 F 13.9
22:00 258.6 F 20.0
24:00 268.8 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.1 F 29.0
2:00 272.2 F 28.6
3:00 271.6 F 28.7
4:00 272.5 F 28.9

Min 271.2 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.3 F 0.0
5:00 216.8 F 0.0
10:00 232.7 F 0.0
15:00 218.7 F 0.0
20:00 210.9 F 0.0
25:00 217.4 F 0.0
30:00 212.2 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:04
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:03:29
CYCLE COMPLETE

06/22/2021 09:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____