
Midmark M11 - v1.0.5
Total Cycles: 2217

Name: RITTER2
Sterilizer ID: V2313228
04/04/2023 02:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.2 F 0.0
2:00 184.8 F 4.1
4:00 202.4 F 9.1
6:00 222.2 F 8.2
8:00 232.0 F 10.9
10:00 233.5 F 8.7
12:00 238.9 F 10.1
14:00 248.1 F 14.6
16:00 260.6 F 21.7
18:00 272.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.4 F 29.4
2:00 271.3 F 28.7
3:00 272.3 F 29.6
4:00 271.9 F 29.0

Min 271.2 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 231.3 F 0.0
10:00 228.4 F 0.0
15:00 226.7 F 0.0
20:00 230.7 F 0.0
25:00 229.2 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 18:32
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:52
CYCLE COMPLETE

04/04/2023 03:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____