

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3010

Name: RITTER 1  
Sterilizer ID: V2178198  
03/16/2023 07:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.7 F 0.0  
2:00 79.4 F 0.0  
4:00 89.0 F 0.0  
6:00 120.2 F 0.4  
8:00 149.0 F 1.9  
10:00 177.2 F 4.0  
12:00 201.0 F 7.1  
14:00 214.8 F 10.6  
16:00 224.5 F 7.3  
18:00 226.2 F 8.0  
20:00 236.7 F 12.2  
22:00 234.3 F 8.6  
24:00 242.2 F 11.3  
26:00 244.2 F 12.4  
28:00 255.8 F 18.5  
30:00 266.3 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 271.8 F 28.8  
2:00 272.3 F 28.8  
3:00 271.4 F 28.3  
4:00 272.5 F 29.2  
  
Min 271.4 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.4 F 0.0  
5:00 201.9 F 0.0  
10:00 208.9 F 0.0  
15:00 216.6 F 0.0  
20:00 226.7 F 0.0  
25:00 229.7 F 0.0  
30:00 223.7 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 31:33  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:09:33  
CYCLE COMPLETE

03/16/2023 08:44 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_