

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007590  
7:50 25/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CYCLE ABORTED 2:49  
CASSETTE REMOVED  
NOT STERILE  
END TIME 2:49  
-----

Digital Signature #  
00920B6CA869DE0B  
-----