
Midmark M11 - v1.0.0
Total Cycles: 980

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2021 06:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.0 F 0.0
2:00 76.7 F 0.0
4:00 82.0 F 0.0
6:00 118.7 F 0.0
8:00 149.5 F 0.1
10:00 175.8 F 0.3
12:00 196.4 F 2.4
14:00 211.6 F 5.3
16:00 219.5 F 8.4
18:00 228.9 F 11.9
20:00 228.4 F 6.8
22:00 239.0 F 11.5
24:00 235.8 F 8.7
26:00 241.7 F 10.9
28:00 246.2 F 13.1
30:00 257.7 F 19.2
32:00 267.9 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.0
1:00 272.1 F 28.6
2:00 271.9 F 28.1
3:00 271.9 F 28.5
4:00 272.2 F 28.4

Min 271.3 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 209.3 F 0.0
10:00 210.5 F 0.0
15:00 217.0 F 0.0
20:00 226.6 F 0.0
25:00 228.4 F 0.0
30:00 225.7 F 0.0

Drying Complete

FILLING: 1:01
HEATING: 33:14
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:11:26
CYCLE COMPLETE

07/08/2021 07:57 AM
ACCEPTED: X
REJECTED:
Operator: U02