

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1949

Name: RITTER2  
Sterilizer ID: V2313228  
01/19/2023 10:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.7 F 0.0  
2:00 199.5 F 3.4  
4:00 194.5 F 8.4  
6:00 216.3 F 9.5  
8:00 229.8 F 9.6  
10:00 233.2 F 8.6  
12:00 242.8 F 12.4  
14:00 245.0 F 13.1  
16:00 258.0 F 20.2  
18:00 269.6 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.3 F 29.7  
2:00 271.7 F 29.0  
3:00 271.5 F 29.4  
4:00 272.2 F 29.6  
  
Min 271.0 F 28.8  
Max 272.4 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.4 F 0.0  
5:00 225.3 F 0.0  
10:00 233.4 F 0.0  
15:00 231.2 F 0.0  
20:00 227.8 F 0.0  
25:00 228.1 F 0.0  
30:00 223.5 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 19:00  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:53  
CYCLE COMPLETE

01/19/2023 11:50 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_