
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001061
11:09 06/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:28
STERILIZING 3:31
136.0 C 220kPa 3:31
136.1 C 222kPa 4:01
136.2 C 222kPa 4:31
136.3 C 223kPa 5:01
136.4 C 225kPa 5:31
136.3 C 224kPa 6:01
136.4 C 225kPa 6:31
136.3 C 224kPa 7:01
136.3 C 224kPa 7:01
VENTING 7:01
AIR DRYING 7:55
CYCLE COMPLETE 22:55

Digital Signature #
CB528ADC8909E7E7
