
Midmark M11 - v1.0.0
Total Cycles: 4582

Name: RITTER 1
Sterilizer ID: V2178198
05/24/2024 02:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.2 F 0.0
2:00 151.3 F 2.8
4:00 170.2 F 6.4
6:00 190.3 F 10.1
8:00 219.3 F 6.9
10:00 225.2 F 7.7
12:00 238.2 F 12.0
14:00 233.5 F 8.1
16:00 244.6 F 13.4
18:00 242.8 F 11.6
20:00 254.4 F 17.5
22:00 264.9 F 23.8
24:00 271.1 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.6 F 29.1
2:00 272.5 F 28.7
3:00 271.3 F 28.2
4:00 271.3 F 28.6

Min 271.0 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.6 F 0.0
5:00 213.5 F 0.0
10:00 190.6 F 0.0
15:00 189.3 F 0.0
20:00 186.4 F 0.0
25:00 187.3 F 0.0
30:00 187.4 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 24:01
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:44
CYCLE COMPLETE

05/24/2024 03:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____