
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001472
12:11 10/04/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:23
PRESSURIZING 6:13
STERILIZING 7:50
136.0 C 220kPa 7:50
136.5 C 221kPa 8:20
136.4 C 223kPa 8:50
136.5 C 224kPa 9:20
136.2 C 221kPa 9:50
136.4 C 223kPa 10:20
136.5 C 223kPa 10:50
136.4 C 223kPa 11:20
VENTING 11:20
AIR DRYING 12:25
CYCLE COMPLETE 24:58

Digital Signature #
9EC5527DA9BEFFDF
