
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001767
10:54 02/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:59
STERILIZING 4:12
136.4 C 223kPa 4:12
136.4 C 225kPa 4:42
136.4 C 222kPa 5:12
136.4 C 222kPa 5:42
136.4 C 223kPa 6:12
136.3 C 223kPa 6:42
136.3 C 224kPa 7:12
136.3 C 224kPa 7:42
136.3 C 224kPa 7:42
VENTING 7:42
AIR DRYING 8:29
CYCLE COMPLETE 8:35

Digital Signature #
5A47A81FB171FDED
