
Midmark M11 - v1.0.0
Total Cycles: 1103

Name: RITTER 1
Sterilizer ID: V2178198
08/20/2021 07:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	72.9 F	0.0
2:00	80.5 F	0.0
4:00	92.8 F	0.0
6:00	132.3 F	0.0
8:00	162.1 F	0.1
10:00	185.1 F	0.1
12:00	205.2 F	0.3
14:00	213.5 F	3.3
16:00	221.2 F	5.6
18:00	228.5 F	8.6
20:00	238.1 F	12.2
22:00	234.1 F	8.6
24:00	244.3 F	12.4
26:00	243.3 F	11.9
28:00	240.9 F	10.5
30:00	253.2 F	16.7
32:00	263.6 F	22.8
34:00	272.6 F	28.5

STERILIZING

mm:ss	Degrees	PSI
0:00	271.4 F	28.0
1:00	272.5 F	28.8
2:00	272.1 F	28.2
3:00	271.4 F	28.1
4:00	272.6 F	28.8

Min 271.3 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	180.7 F	0.0
	5:00	221.4 F	0.0
	10:00	223.8 F	0.0
	15:00	220.1 F	0.0
	20:00	212.8 F	0.0
Drying Complete	25:00	217.6 F	0.0
	30:00	224.1 F	0.0

FILLING: 0:48
HEATING: 34:20
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:12:26
CYCLE COMPLETE

08/20/2021 08:25 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____