
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 003759
9:04 06/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:51
STERILIZING 4:44
136.0 C 220kPa 4:44
136.3 C 219kPa 5:14
136.4 C 221kPa 5:44
136.2 C 223kPa 6:14
136.4 C 222kPa 6:44
136.3 C 221kPa 7:14
136.4 C 222kPa 7:44
136.2 C 220kPa 8:14
136.2 C 220kPa 8:14
VENTING 8:14
AIR DRYING 9:04
CYCLE COMPLETE 9:25

Digital Signature #
BC624BDCA27CCD16
