
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001202
11:28 03/08/2022
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 4:24
STERILIZING 5:12
136.0 C 220kPa 5:12
136.4 C 225kPa 5:42
136.2 C 223kPa 6:12
136.4 C 222kPa 6:42
136.1 C 221kPa 7:12
136.2 C 223kPa 7:42
136.4 C 224kPa 8:12
136.2 C 222kPa 8:42
136.4 C 223kPa 9:12
136.4 C 224kPa 9:42
136.3 C 223kPa 10:12
136.5 C 225kPa 10:42
136.3 C 223kPa 11:12
136.2 C 223kPa 11:42
136.3 C 224kPa 12:12
136.4 C 224kPa 12:42
136.1 C 219kPa 13:12
136.0 C 222kPa 13:42
136.3 C 222kPa 14:12
136.3 C 223kPa 14:42
136.2 C 223kPa 15:12
136.2 C 223kPa 15:12
VENTING 15:12
AIR DRYING 15:56
CYCLE COMPLETE 30:56

Digital Signature #
3E7A8D072B72C9DB
