
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000134
16:29 08/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:35
PRESSURIZING 4:35
STERILIZING 6:05
136.4 C 226kPa 6:05
136.1 C 222kPa 6:35
136.4 C 224kPa 7:05
136.4 C 225kPa 7:35
136.4 C 226kPa 8:05
136.2 C 224kPa 8:35
136.4 C 226kPa 9:05
136.3 C 224kPa 9:35
136.3 C 224kPa 9:35
VENTING 9:35
AIR DRYING 10:37
CYCLE COMPLETE 12:51

Digital Signature #
7FF1F9AA86DD2800
