
Midmark M11 - v1.0.0
Total Cycles: 4619

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2024 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.3 F 0.0
2:00 101.9 F 0.9
4:00 125.1 F 3.0
6:00 148.6 F 5.6
8:00 171.9 F 8.4
10:00 193.1 F 12.0
12:00 208.1 F 3.5
14:00 220.0 F 6.7
16:00 228.3 F 10.2
18:00 231.9 F 8.5
20:00 237.4 F 10.6
22:00 237.8 F 8.6
24:00 246.4 F 13.7
26:00 257.0 F 19.4
28:00 266.8 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.4 F 29.4
2:00 272.2 F 29.0
3:00 271.2 F 28.5
4:00 272.0 F 29.2

Min 271.1 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 225.0 F 0.0
10:00 219.3 F 0.0
15:00 228.4 F 0.0
20:00 238.6 F 0.0
25:00 231.4 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 29:42
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:08:29
CYCLE COMPLETE

06/07/2024 03:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____