
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001

Unit # : 000

WATER QUALITY

0.3uS / 0.1ppm

CYCLE NUMBER 001077

12:19 11/01/2021

UNWRAPPED

135 C for 3.5 min.

CYCLE START 0:00

CONDITIONING 3:36

PRESSURIZING 3:37

STERILIZING 4:42

136.4 C 222kPa 4:42

136.3 C 224kPa 5:12

136.3 C 223kPa 5:42

136.1 C 223kPa 6:12

136.4 C 224kPa 6:42

136.3 C 224kPa 7:12

136.2 C 223kPa 7:42

136.4 C 225kPa 8:12

136.4 C 225kPa 8:12

VENTING 8:12

AIR DRYING 9:05

CYCLE COMPLETE 9:06

Digital Signature #

3732F9634DCEED61
