

-----  
Midmark M11 - v1.0.0  
Total Cycles: 972

Name: RITTER 1  
Sterilizer ID: V2178198  
07/07/2021 06:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.3 F 0.0  
2:00 82.1 F 0.0  
4:00 80.1 F 0.0  
6:00 105.2 F 0.1  
8:00 136.2 F 0.2  
10:00 169.4 F 0.5  
12:00 191.9 F 3.4  
14:00 219.9 F 6.3  
16:00 225.1 F 9.8  
18:00 235.8 F 13.5  
20:00 231.7 F 8.3  
22:00 242.1 F 12.8  
24:00 239.4 F 10.2  
26:00 236.6 F 8.4  
28:00 249.4 F 14.7  
30:00 261.0 F 21.1  
32:00 271.5 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 27.9  
1:00 272.5 F 28.6  
2:00 271.3 F 27.8  
3:00 272.5 F 28.7  
4:00 271.7 F 27.9  
  
Min 271.2 F 27.8  
Max 272.6 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.6 F 0.0  
5:00 213.7 F 0.0  
10:00 175.8 F 0.0  
15:00 210.2 F 0.0  
20:00 161.0 F 0.0  
25:00 154.6 F 0.0  
30:00 150.5 F 0.0  
  
Drying Complete

FILLING: 0:46  
HEATING: 32:39  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:10:35  
CYCLE COMPLETE

07/07/2021 08:11 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01