
Midmark M11 - v1.0.0
Total Cycles: 1125

Name: RITTER 1
Sterilizer ID: V2178198
08/27/2021 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.0 F 0.0
2:00 174.0 F 2.4
4:00 193.5 F 6.0
6:00 210.9 F 10.4
8:00 224.2 F 6.7
10:00 231.8 F 10.1
12:00 235.3 F 8.5
14:00 244.4 F 13.2
16:00 243.5 F 11.9
18:00 255.9 F 18.5
20:00 266.9 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.5 F 28.9
2:00 271.3 F 28.3
3:00 272.5 F 29.1
4:00 271.8 F 28.5

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.9 F 0.0
5:00 217.3 F 0.0
10:00 215.5 F 0.0
15:00 214.2 F 0.0
20:00 216.3 F 0.0
25:00 222.5 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 21:30
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

08/27/2021 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____