
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000562
13:46 04/08/2021
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 5:33
STERILIZING 6:34
136.0 C 226kPa 6:34
136.4 C 232kPa 7:04
136.4 C 230kPa 7:34
136.4 C 230kPa 8:04
136.1 C 230kPa 8:34
135.9 C 230kPa 9:04
136.3 C 230kPa 9:34
136.3 C 230kPa 10:04
136.4 C 233kPa 10:34
136.1 C 230kPa 11:04
136.4 C 231kPa 11:34
136.3 C 231kPa 12:04
136.4 C 232kPa 12:34
136.2 C 231kPa 13:04
136.1 C 229kPa 13:34
136.4 C 231kPa 14:04
136.1 C 227kPa 14:34
136.4 C 233kPa 15:04
136.4 C 233kPa 15:34
136.1 C 230kPa 16:04
136.4 C 233kPa 16:34
136.4 C 233kPa 16:34
VENTING 16:34
AIR DRYING 17:28
CYCLE COMPLETE 32:28

Digital Signature #
03D89F53AFE96668
