

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001151  
10:02 05/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:37  
PRESSURIZING 3:43  
STERILIZING 4:54  
136.0 C 222kPa 4:54  
136.3 C 223kPa 5:25  
136.2 C 222kPa 5:55  
136.3 C 224kPa 6:25  
136.1 C 222kPa 6:55  
136.4 C 220kPa 7:25  
136.4 C 225kPa 7:55  
136.3 C 224kPa 8:25  
136.3 C 224kPa 8:25  
VENTING 8:25  
AIR DRYING 9:18  
CASSETTE REMOVED  
CYCLE COMPLETE 10:25  
-----

Digital Signature #  
D8042CD2FF96BE26  
-----