
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000503
8:33 22/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:34
STERILIZING 4:36
136.0 C 220kPa 4:36
136.3 C 220kPa 5:06
136.4 C 224kPa 5:36
136.3 C 224kPa 6:06
136.2 C 224kPa 6:36
136.4 C 223kPa 7:06
136.4 C 225kPa 7:36
136.3 C 222kPa 8:06
136.3 C 222kPa 8:06
VENTING 8:06
AIR DRYING 9:00
CYCLE COMPLETE 9:05

Digital Signature #
8A56292B6FA3760F
