

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4550

Name: RITTER 1  
Sterilizer ID: V2178198  
05/14/2024 09:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
2:00 168.9 F 2.0  
4:00 175.5 F 5.3  
6:00 194.6 F 9.1  
8:00 210.1 F 13.3  
10:00 218.3 F 6.3  
12:00 230.3 F 10.6  
14:00 232.1 F 7.7  
16:00 242.6 F 12.2  
18:00 240.3 F 10.4  
20:00 252.3 F 16.3  
22:00 262.9 F 22.4  
24:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.0 F 28.5  
2:00 272.6 F 28.8  
3:00 272.5 F 29.1  
4:00 271.8 F 29.0

Min 271.5 F 28.5  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
5:00 234.0 F 0.0  
10:00 228.8 F 0.0  
15:00 234.1 F 0.0  
20:00 232.0 F 0.0  
25:00 234.8 F 0.0  
30:00 229.4 F 0.0  
Drying Complete

FILLING: 1:21  
HEATING: 25:04  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:04:10  
CYCLE COMPLETE

05/14/2024 10:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_