
Midmark M11 - v1.0.0
Total Cycles: 5262

Name: RITTER 1
Sterilizer ID: V2178198
01/22/2025 05:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.9 F 0.0
2:00 133.3 F 1.5
4:00 150.1 F 4.7
6:00 178.0 F 8.0
8:00 195.1 F 11.7
10:00 202.5 F 3.7
12:00 216.8 F 8.0
14:00 229.9 F 12.2
16:00 230.1 F 8.2
18:00 241.8 F 13.4
20:00 240.9 F 11.2
22:00 252.7 F 17.0
24:00 263.6 F 23.4
26:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 271.9 F 29.2
2:00 272.6 F 29.3
3:00 272.1 F 28.9
4:00 271.4 F 28.5

Min 271.2 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 199.6 F 0.0
10:00 195.3 F 0.0
15:00 186.1 F 0.0
20:00 188.5 F 0.0
25:00 199.2 F 0.0
30:00 206.8 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 26:19
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:05:47
CYCLE COMPLETE

01/22/2025 06:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____