

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001250  
14:22 07/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:59  
PRESSURIZING 4:00  
STERILIZING 5:34  
136.5 C 225kPa 5:34  
136.4 C 224kPa 6:04  
136.3 C 223kPa 6:34  
136.1 C 221kPa 7:04  
136.3 C 223kPa 7:34  
136.4 C 225kPa 8:04  
136.4 C 221kPa 8:34  
136.1 C 222kPa 9:04  
136.1 C 222kPa 9:04  
VENTING 9:04  
AIR DRYING 9:58  
CYCLE COMPLETE 22:31  
-----

Digital Signature #  
712FFD19B42781F1  
-----