
Midmark M11 - v1.0.5
Total Cycles: 477

Name: RITTER2
Sterilizer ID: V2313228
10/25/2021 06:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 83.9 F 0.0
4:00 132.1 F 0.1
6:00 100.0 F 0.3
8:00 123.3 F 0.3
10:00 156.6 F 0.8
12:00 184.4 F 3.9
14:00 221.9 F 7.7
16:00 223.0 F 11.4
18:00 228.3 F 7.0
20:00 231.1 F 12.0
22:00 236.3 F 9.5
24:00 237.9 F 9.8
26:00 247.3 F 14.2
28:00 258.6 F 20.3
30:00 268.9 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 271.8 F 29.5
2:00 272.4 F 30.0
3:00 271.3 F 29.2
4:00 272.8 F 29.7

Min 271.2 F 29.1
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 226.3 F 0.0
10:00 208.3 F 0.0
15:00 217.7 F 0.0
20:00 216.3 F 0.0
25:00 214.7 F 0.0
30:00 169.2 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:53
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:47
CYCLE COMPLETE

10/25/2021 08:14 AM
ACCEPTED: X
REJECTED:
Operator: U01