
Midmark M11 - v1.0.5
Total Cycles: 2859

Name: RITTER2
Sterilizer ID: V2313228
09/11/2023 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.2 F 0.0
2:00 184.3 F 2.4
4:00 190.6 F 6.6
6:00 211.1 F 11.2
8:00 224.6 F 5.8
10:00 233.4 F 10.5
12:00 237.0 F 8.3
14:00 244.2 F 13.4
16:00 243.8 F 12.5
18:00 255.6 F 18.7
20:00 266.3 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.7 F 29.4
2:00 271.5 F 28.7
3:00 272.6 F 29.6
4:00 271.9 F 29.0

Min 271.5 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 210.1 F 0.0
10:00 218.7 F 0.0
15:00 223.4 F 0.0
20:00 223.6 F 0.0
25:00 223.2 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 21:42
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:02
CYCLE COMPLETE

09/11/2023 11:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____