
Midmark M11 - v1.0.0
Total Cycles: 1689

Name: RITTER 1
Sterilizer ID: V2178198
03/02/2022 07:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.0 F 0.0
2:00 77.0 F 0.0
4:00 83.8 F 0.2
6:00 111.1 F 1.1
8:00 141.1 F 2.4
10:00 168.1 F 4.5
12:00 190.1 F 7.7
14:00 204.5 F 11.1
16:00 216.6 F 5.1
18:00 223.3 F 7.9
20:00 234.6 F 12.1
22:00 233.4 F 8.4
24:00 242.8 F 12.0
26:00 243.3 F 12.0
28:00 254.9 F 18.1
30:00 265.7 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.4 F 29.0
2:00 271.7 F 28.3
3:00 271.6 F 28.6
4:00 272.4 F 29.0

Min 271.2 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.6 F 0.0
5:00 200.2 F 0.0
10:00 209.8 F 0.0
15:00 214.4 F 0.0
20:00 213.2 F 0.0
25:00 213.9 F 0.0
30:00 217.9 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:50
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:44
CYCLE COMPLETE

03/02/2022 09:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____