

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1922

Name: RITTER2  
Sterilizer ID: V2313228  
01/11/2023 03:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.6 F 0.0  
2:00 193.1 F 3.9  
4:00 198.4 F 9.4  
6:00 223.2 F 7.4  
8:00 232.3 F 10.9  
10:00 235.0 F 8.9  
12:00 239.0 F 9.6  
14:00 249.4 F 15.4  
16:00 262.0 F 22.7  
18:00 272.8 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.6 F 29.8  
2:00 272.0 F 29.3  
3:00 271.2 F 29.0  
4:00 272.3 F 29.9  
  
Min 271.2 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.7 F 0.0  
5:00 233.3 F 0.0  
10:00 227.2 F 0.0  
15:00 234.5 F 0.0  
20:00 230.8 F 0.0  
25:00 226.8 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 2:07  
HEATING: 18:21  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:38  
CYCLE COMPLETE

01/11/2023 04:06 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_