
Midmark M11 - v1.0.0
Total Cycles: 1687

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2022 01:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.4 F 0.0
2:00 140.8 F 1.3
4:00 180.7 F 3.9
6:00 195.1 F 7.8
8:00 220.7 F 12.7
10:00 226.5 F 6.7
12:00 239.0 F 11.7
14:00 235.9 F 8.8
16:00 239.4 F 10.0
18:00 247.3 F 13.7
20:00 259.5 F 20.4
22:00 270.4 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 271.9 F 28.7
2:00 272.3 F 28.6
3:00 271.0 F 28.0
4:00 272.4 F 29.0

Min 271.0 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 238.1 F 0.0
10:00 223.2 F 0.0
15:00 227.1 F 0.0
20:00 226.9 F 0.0
25:00 225.5 F 0.0
30:00 231.3 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 22:42
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:57
CYCLE COMPLETE

03/01/2022 03:21 PM
ACCEPTED: X
REJECTED:
Operator: U02