
Midmark M11 - v1.0.0
Total Cycles: 824

Name: RITTER 1
Sterilizer ID: V2178198
05/19/2021 02:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.1 F 0.0
2:00 150.6 F 1.8
4:00 174.6 F 4.6
6:00 195.3 F 8.8
8:00 210.5 F 11.3
10:00 225.8 F 7.3
12:00 237.2 F 12.4
14:00 236.0 F 9.2
16:00 239.7 F 9.7
18:00 247.0 F 13.5
20:00 258.9 F 20.0
22:00 269.9 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.6 F 28.8
2:00 271.6 F 28.2
3:00 272.4 F 28.5
4:00 271.8 F 28.4

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.6 F 0.0
5:00 234.3 F 0.0
10:00 240.3 F 0.0
15:00 208.9 F 0.0
20:00 211.8 F 0.0
25:00 211.5 F 0.0
30:00 212.6 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 22:48
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:13
CYCLE COMPLETE

05/19/2021 03:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____