

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3589

Name: RITTER2  
Sterilizer ID: V2313228  
05/03/2024 01:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.2 F 0.0  
2:00 134.0 F 1.8  
4:00 153.9 F 4.7  
6:00 177.9 F 8.2  
8:00 195.9 F 12.1  
10:00 212.2 F 5.2  
12:00 221.4 F 9.2  
14:00 233.5 F 13.5  
16:00 234.3 F 10.1  
18:00 236.1 F 9.3  
20:00 245.3 F 13.7  
22:00 256.0 F 19.4  
24:00 265.8 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.3  
1:00 271.4 F 29.4  
2:00 272.4 F 30.2  
3:00 272.4 F 30.0  
4:00 271.8 F 29.6

Min 271.1 F 29.2  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.2 F 0.0  
5:00 209.0 F 0.0  
10:00 200.6 F 0.0  
15:00 194.4 F 0.0  
20:00 191.5 F 0.0  
25:00 194.7 F 0.0  
30:00 206.5 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 25:46  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:04:36  
CYCLE COMPLETE

05/03/2024 02:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_