
Midmark M11 - v1.0.5
Total Cycles: 905

Name: RITTER2
Sterilizer ID: V2313228
03/25/2022 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.5 F 0.0
2:00 157.6 F 0.2
4:00 172.8 F 1.9
6:00 199.5 F 6.8
8:00 214.4 F 12.2
10:00 228.0 F 7.6
12:00 239.7 F 13.0
14:00 239.3 F 11.4
16:00 239.4 F 10.6
18:00 253.2 F 17.5
20:00 265.2 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.6 F 29.4
2:00 271.2 F 29.1
3:00 272.1 F 29.4
4:00 271.1 F 29.0

Min 271.1 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 241.1 F 0.0
10:00 212.3 F 0.0
15:00 199.5 F 0.0
20:00 201.4 F 0.0
25:00 203.7 F 0.0
30:00 201.7 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:03
CYCLE COMPLETE

03/25/2022 09:47 AM
ACCEPTED: X
REJECTED:
Operator: U02