
Midmark M11 - v1.0.0
Total Cycles: 5287

Name: RITTER 1
Sterilizer ID: V2178198
01/28/2025 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.9 F 0.0
2:00 159.5 F 3.7
4:00 184.7 F 8.1
6:00 206.0 F 12.9
8:00 221.1 F 6.1
10:00 231.0 F 11.2
12:00 231.5 F 7.4
14:00 245.4 F 13.4
16:00 243.0 F 11.8
18:00 255.7 F 18.4
20:00 266.9 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.6 F 28.9
2:00 271.4 F 28.2
3:00 272.4 F 29.1
4:00 271.9 F 28.5

Min 271.3 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.4 F 0.0
5:00 223.9 F 0.0
10:00 211.2 F 0.0
15:00 215.3 F 0.0
20:00 221.3 F 0.0
25:00 219.5 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

01/28/2025 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____