
Midmark M11 - v1.0.0
Total Cycles: 2841

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2023 11:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.2 F 0.0
2:00 204.7 F 4.2
4:00 202.8 F 9.6
6:00 220.5 F 6.2
8:00 230.7 F 10.1
10:00 237.4 F 8.1
12:00 244.8 F 13.3
14:00 243.8 F 12.0
16:00 256.7 F 18.8
18:00 268.0 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.5 F 28.9
2:00 271.5 F 28.1
3:00 272.4 F 29.1
4:00 272.1 F 28.5

Min 271.4 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.3 F 0.0
5:00 226.7 F 0.0
10:00 233.4 F 0.0
15:00 238.5 F 0.0
20:00 234.7 F 0.0
25:00 224.6 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

02/01/2023 12:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____