

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1246

Name: RITTER2  
Sterilizer ID: V2313228  
07/04/2022 07:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.4 F 0.0  
2:00 74.2 F 0.0  
4:00 79.6 F 0.0  
6:00 103.2 F 0.1  
8:00 131.0 F 0.4  
10:00 162.7 F 2.8  
12:00 183.2 F 5.6  
14:00 205.1 F 9.2  
16:00 217.9 F 13.5  
18:00 225.2 F 7.7  
20:00 235.2 F 12.1  
22:00 233.9 F 9.1  
24:00 239.9 F 10.4  
26:00 245.5 F 13.4  
28:00 256.8 F 19.4  
30:00 266.8 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.2 F 29.3  
2:00 271.7 F 29.3  
3:00 271.9 F 29.1  
4:00 272.1 F 29.6

Min 271.2 F 28.8  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.0 F 0.0  
5:00 198.3 F 0.0  
10:00 205.4 F 0.0  
15:00 210.1 F 0.0  
20:00 212.6 F 0.0  
25:00 212.1 F 0.0  
30:00 212.1 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:34  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:09:31  
CYCLE COMPLETE

07/04/2022 08:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_