
Midmark M11 - v1.0.0
Total Cycles: 852

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2021 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.1 F 0.0
2:00 182.3 F 2.8
4:00 212.5 F 7.6
6:00 218.2 F 13.3
8:00 229.4 F 8.0
10:00 241.5 F 13.4
12:00 241.0 F 11.1
14:00 239.0 F 9.7
16:00 253.1 F 16.6
18:00 264.9 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.4 F 28.4
2:00 272.0 F 28.3
3:00 272.2 F 28.3
4:00 272.0 F 28.4

Min 271.5 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 224.7 F 0.0
10:00 233.1 F 0.0
15:00 229.0 F 0.0
20:00 189.4 F 0.0
25:00 194.1 F 0.0
30:00 192.1 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:50
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

05/31/2021 12:49 PM
ACCEPTED: X
REJECTED:
Operator: U01