
Midmark M11 - v1.0.5
Total Cycles: 3155

Name: RITTER2
Sterilizer ID: V2313228
11/22/2023 03:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.5 F 0.0
2:00 93.8 F 0.0
4:00 110.0 F 1.1
6:00 143.8 F 4.0
8:00 171.8 F 7.1
10:00 193.4 F 10.7
12:00 211.0 F 5.2
14:00 221.7 F 7.7
16:00 232.6 F 12.2
18:00 232.7 F 8.9
20:00 239.8 F 10.9
22:00 244.4 F 13.1
24:00 256.3 F 19.4
26:00 267.1 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.6 F 29.8
2:00 271.3 F 29.0
3:00 272.3 F 30.0
4:00 272.0 F 29.5

Min 271.0 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.4 F 0.0
5:00 223.1 F 0.0
10:00 192.5 F 0.0
15:00 193.2 F 0.0
20:00 193.7 F 0.0
25:00 193.1 F 0.0
30:00 193.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 27:29
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:05:46
CYCLE COMPLETE

11/22/2023 04:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____