
Midmark M11 - v1.0.0
Total Cycles: 1617

Name: RITTER 1
Sterilizer ID: V2178198
02/10/2022 12:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.9 F 0.0
2:00 198.3 F 4.8
4:00 224.7 F 9.8
6:00 226.1 F 6.4
8:00 236.8 F 11.0
10:00 234.8 F 8.6
12:00 239.4 F 10.1
14:00 247.8 F 14.2
16:00 260.4 F 21.1
18:00 271.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.3 F 28.8
2:00 272.0 F 29.1
3:00 271.9 F 28.6
4:00 271.9 F 29.0

Min 271.2 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 231.3 F 0.0
10:00 218.9 F 0.0
15:00 210.9 F 0.0
20:00 218.4 F 0.0
25:00 225.6 F 0.0
30:00 201.1 F 0.0
Drying Complete

FILLING: 2:33
HEATING: 18:34
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:26
CYCLE COMPLETE

02/10/2022 01:53 PM
ACCEPTED: X
REJECTED:
Operator: U02