
Midmark M11 - v1.0.0
Total Cycles: 2047

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2022 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.5 F 0.0
2:00 182.1 F 3.7
4:00 185.6 F 7.6
6:00 212.8 F 12.3
8:00 221.2 F 6.4
10:00 233.7 F 11.4
12:00 233.2 F 8.2
14:00 241.5 F 10.9
16:00 245.5 F 13.1
18:00 258.1 F 19.8
20:00 268.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.3
2:00 272.2 F 28.8
3:00 271.3 F 28.3
4:00 271.9 F 29.0

Min 271.2 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 228.6 F 0.0
10:00 229.2 F 0.0
15:00 232.4 F 0.0
20:00 227.1 F 0.0
25:00 233.7 F 0.0
30:00 232.3 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

06/22/2022 02:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____