

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4287

Name: RITTER 1  
Sterilizer ID: V2178198  
02/09/2024 09:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.4 F 0.0  
2:00 75.0 F 0.2  
4:00 80.4 F 1.2  
6:00 106.9 F 3.2  
8:00 135.3 F 5.9  
10:00 163.0 F 9.1  
12:00 182.6 F 12.5  
14:00 204.9 F 3.0  
16:00 213.5 F 6.1  
18:00 222.0 F 9.5  
20:00 232.7 F 13.5  
22:00 234.5 F 9.4  
24:00 239.1 F 10.3  
26:00 243.7 F 12.4  
28:00 255.3 F 18.4  
30:00 265.7 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.5 F 29.1  
2:00 271.5 F 28.8  
3:00 272.2 F 28.8  
4:00 272.0 F 29.1

Min 271.2 F 28.4  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.7 F 0.0  
5:00 233.6 F 0.0  
10:00 222.2 F 0.0  
15:00 225.1 F 0.0  
20:00 235.0 F 0.0  
25:00 225.8 F 0.0  
30:00 222.6 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:47  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:10:00  
CYCLE COMPLETE

02/09/2024 10:19 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_