
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000970
10:52 09/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:09
PRESSURIZING 2:14
STERILIZING 3:16
136.4 C 228kPa 3:16
136.3 C 232kPa 3:46
136.4 C 229kPa 4:16
136.3 C 230kPa 4:46
136.1 C 230kPa 5:16
136.4 C 232kPa 5:46
136.2 C 231kPa 6:16
136.4 C 232kPa 6:46
136.4 C 232kPa 6:46
VENTING 6:46
AIR DRYING 7:37
CYCLE COMPLETE 7:42

Digital Signature #
FE9A70B067D7F4C3
