
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000798
7:15 04/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:21
PRESSURIZING 1:30
STERILIZING 2:33
136.4 C 231kPa 2:33
136.3 C 232kPa 3:03
136.4 C 229kPa 3:33
136.4 C 231kPa 4:03
136.4 C 232kPa 4:33
136.2 C 230kPa 5:03
136.4 C 232kPa 5:33
136.4 C 232kPa 6:03
136.4 C 232kPa 6:03
VENTING 6:03
AIR DRYING 6:54
CYCLE COMPLETE 7:17

Digital Signature #
4C5F4C0272D8BAB2
