

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000186  
12:31 17/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:36  
PRESSURIZING 4:36  
STERILIZING 5:39  
136.0 C 221kPa 5:39  
136.2 C 224kPa 6:09  
136.4 C 223kPa 6:39  
136.3 C 225kPa 7:09  
136.4 C 225kPa 7:39  
136.3 C 225kPa 8:09  
136.4 C 226kPa 8:39  
136.4 C 226kPa 9:09  
136.4 C 226kPa 9:09  
VENTING 9:09  
AIR DRYING 10:09  
CYCLE COMPLETE 25:09  
-----

Digital Signature #  
1E771A1AAA95AAFD  
-----