
Midmark M11 - v1.0.5
Total Cycles: 2293

Name: RITTER2
Sterilizer ID: V2313228
04/24/2023 02:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 213.0 F 0.0
2:00 207.5 F 4.5
4:00 211.0 F 10.7
6:00 226.2 F 7.3
8:00 239.0 F 12.3
10:00 242.8 F 12.6
12:00 243.5 F 12.3
14:00 257.6 F 19.8
16:00 269.7 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.6 F 29.6
2:00 271.4 F 28.9
3:00 272.5 F 29.8
4:00 272.2 F 29.4

Min 271.3 F 28.8
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 226.0 F 0.0
10:00 232.4 F 0.0
15:00 228.4 F 0.0
20:00 229.3 F 0.0
25:00 230.9 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 2:47
HEATING: 16:56
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:56:58
CYCLE COMPLETE

04/24/2023 03:25 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____