

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002305  
9:51 20/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:59  
PRESSURIZING 2:06  
STERILIZING 3:14  
136.0 C 221kPa 3:14  
136.4 C 225kPa 3:44  
136.4 C 222kPa 4:14  
136.3 C 224kPa 4:44  
136.4 C 224kPa 5:14  
136.1 C 222kPa 5:44  
136.3 C 223kPa 6:14  
136.4 C 224kPa 6:44  
136.4 C 224kPa 6:44  
VENTING 6:44  
AIR DRYING 7:39  
CYCLE COMPLETE 8:35  
-----

Digital Signature #  
13BC02D6E1AC9185  
-----