

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3824

Name: RITTER2  
Sterilizer ID: V2313228  
07/31/2024 11:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 87.5 F 0.0  
2:00 88.1 F 0.0  
4:00 102.8 F 0.1  
6:00 126.1 F 0.3  
8:00 161.9 F 2.6  
10:00 174.8 F 5.6  
12:00 191.8 F 8.7  
14:00 212.4 F 12.5  
16:00 217.6 F 6.2  
18:00 226.1 F 10.3  
20:00 232.4 F 9.2  
22:00 239.3 F 11.9  
24:00 238.0 F 10.2  
26:00 250.0 F 16.0  
28:00 260.4 F 21.9  
30:00 270.2 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.7 F 30.1  
2:00 272.4 F 29.6  
3:00 271.3 F 29.1  
4:00 272.2 F 30.0

Min 271.3 F 29.1  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.6 F 0.0  
5:00 204.9 F 0.0  
10:00 212.7 F 0.0  
15:00 214.4 F 0.0  
20:00 222.6 F 0.0  
25:00 229.2 F 0.0  
30:00 229.0 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 31:01  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:09:44  
CYCLE COMPLETE

07/31/2024 12:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_