

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000045  
9:45 27/02/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:26  
PRESSURIZING 1:42  
STERILIZING 2:53  
136.0 C 220kPa 2:53  
136.0 C 221kPa 3:23  
136.1 C 223kPa 3:53  
136.4 C 224kPa 4:23  
136.4 C 225kPa 4:53  
136.4 C 222kPa 5:23  
136.4 C 225kPa 5:53  
136.3 C 224kPa 6:23  
136.3 C 224kPa 6:23  
VENTING 6:23  
AIR DRYING 7:15  
CYCLE COMPLETE 8:50  
-----

Digital Signature #  
5D6EE860F3710E6B  
-----