
Midmark M11 - v1.0.0
Total Cycles: 689

Name: MIDMARK
Sterilizer ID: V2178198
02/23/2021 07:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.7 F 0.0
2:00 77.1 F 0.0
4:00 87.6 F 0.0
6:00 115.3 F 0.1
8:00 154.6 F 0.3
10:00 187.2 F 2.3
12:00 209.6 F 5.2
14:00 219.0 F 8.5
16:00 226.2 F 12.3
18:00 229.6 F 6.8
20:00 240.7 F 11.6
22:00 235.6 F 8.5
24:00 241.3 F 10.6
26:00 246.2 F 13.2
28:00 258.3 F 19.7
30:00 269.1 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 28.6
2:00 271.9 F 28.6
3:00 272.2 F 28.4
4:00 272.1 F 28.8

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.6 F 0.0
5:00 236.4 F 0.0
10:00 221.2 F 0.0
15:00 223.4 F 0.0
20:00 228.8 F 0.0
25:00 229.0 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:05
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:08:49
CYCLE COMPLETE

02/23/2021 08:26 AM
ACCEPTED: X
REJECTED:
Operator: U01