

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5023

Name: RITTER 1  
Sterilizer ID: V2178198  
10/25/2024 08:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.9 F 0.0  
2:00 74.2 F 0.0  
4:00 103.8 F 0.0  
6:00 119.5 F 0.1  
8:00 147.3 F 0.4  
10:00 179.1 F 2.1  
12:00 198.4 F 5.4  
14:00 208.6 F 9.2  
16:00 220.0 F 13.0  
18:00 224.3 F 6.6  
20:00 235.5 F 11.0  
22:00 232.3 F 7.0  
24:00 245.7 F 13.3  
26:00 243.5 F 11.8  
28:00 255.9 F 18.2  
30:00 267.0 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.6 F 28.5  
2:00 271.6 F 28.1  
3:00 272.4 F 28.4  
4:00 271.5 F 28.0  
  
Min 271.5 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.0 F 0.0  
5:00 187.8 F 0.0  
10:00 186.2 F 0.0  
15:00 186.2 F 0.0  
20:00 196.6 F 0.0  
25:00 201.6 F 0.0  
30:00 204.4 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 31:28  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:09:32  
CYCLE COMPLETE

10/25/2024 09:28 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_